

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00485 Calib. Date: 05/31/2022 Calib. No.: 00037
Certification File No.: 00486 Cert. Date: 05/31/2022 Cert. No.: 00033
Linearity File No.: 00487 Lin. Date: 05/31/2022 Lin. No.: 00035
Solution File No.: 00494 Soln. Date: 06/30/2022 Soln. No.: 00182
Sequential File No.: 00494 File Date: 06/30/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0096
Control Solution %: 0.100% Expires: 02/17/2024
Solution Control Lot: 22070 Bottle No.: 0303

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:03D		
Control 1 EC	0.104%	15:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 2 EC	0.103%	15:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 3 EC	0.103%	15:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DAVIS

First Name: IAN

MI: G

Signature: _____



Badge No.: 4934

Date: 06/30/2022