

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00304	Calib. Date: 10/24/2016
Certification File No.:	00305	Calib. No.: 00023
Linearity File No.:	00306	Cert. Date: 10/24/2016
Solution File No.:	00310	Cert. No.: 00020
Sequential File No.:	00310	Lin. Date: 10/24/2016
		Lin. No.: 00022
		Soln. Date: 01/02/2017
		Soln. No.: 00113
		File Date: 01/02/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0097
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0528

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:28S		
Control 1 EC	0.100%	20:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30S		
Control 2 EC	0.099%	20:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31S		
Control 3 EC	0.099%	20:32S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 01/02/2017