

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	Calib. No.:	00027
Calibration File No.:	00352	Cert. No.:	00024
Certification File No.:	00353	Lin. No.:	00026
Linearity File No.:	00354	Soln. No.:	00137
Solution File No.:	00363	Calib. Date:	08/23/2018
Sequential File No.:	00363	Cert. Date:	08/23/2018
		Lin. Date:	08/23/2018
		Soln. Date:	01/06/2019
		File Date:	01/06/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0096
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0284

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:08S		
Control 1 EC	0.101%	23:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10S		
Control 2 EC	0.100%	23:10S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:11S		
Control 3 EC	0.100%	23:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 01/06/2019