

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00326	Calib. Date: 09/18/2017
Certification File No.:	00327	Calib. No.: 00025
Linearity File No.:	00328	Cert. Date: 09/18/2017
Solution File No.:	00334	Cert. No.: 00022
Sequential File No.:	00334	Lin. Date: 09/18/2017
		Lin. No.: 00024
		Soln. Date: 01/14/2018
		Soln. No.: 00124
		File Date: 01/14/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0097
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1254

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:38S		
Control 1 EC	0.100%	20:39S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40S		
Control 2 EC	0.097%	20:40S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41S		
Control 3 EC	0.099%	20:42S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 01/14/2018