

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00304	Calib. Date:	10/24/2016
Certification File No.:	00305	Calib. No.:	00023
Linearity File No.:	00306	Cert. Date:	10/24/2016
Solution File No.:	00312	Cert. No.:	00020
Sequential File No.:	00312	Lin. Date:	10/24/2016
		Lin. No.:	00022
		Soln. Date:	02/07/2017
		Soln. No.:	00114
		File Date:	02/07/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0097
Solution Control Lot:	16140	Expires:	06/07/2018
		Bottle No.:	1459

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:55S		
Control 1 EC	0.100%	20:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:56S		
Control 2 EC	0.097%	20:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:58S		
Control 3 EC	0.098%	20:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 02/07/2017