

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00326 Calib. Date: 09/18/2017 Calib. No.: 00025
Certification File No.: 00327 Cert. Date: 09/18/2017 Cert. No.: 00022
Linearity File No.: 00328 Lin. Date: 09/18/2017 Lin. No.: 00024
Solution File No.: 00335 Soln. Date: 02/08/2018 Soln. No.: 00125
Sequential File No.: 00335 File Date: 02/08/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0097
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 0391

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:29S		
Control 1 EC	0.100%	15:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30S		
Control 2 EC	0.097%	15:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		
Control 3 EC	0.098%	15:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 02/08/2018