

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00364	Calib. No.: 00028
Certification File No.:	00365	Cert. No.: 00025
Linearity File No.:	00366	Lin. No.: 00027
Solution File No.:	00368	Soln. No.: 00139
Sequential File No.:	00368	File Date: 02/28/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0097
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0169

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:47S		
Control 1 EC	0.101%	20:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48S		
Control 2 EC	0.098%	20:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:50S		
Control 3 EC	0.098%	20:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 02/28/2019