

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00313 Calib. Date: 04/03/2017 Calib. No.: 00024
Certification File No.: 00314 Cert. Date: 04/03/2017 Cert. No.: 00021
Linearity File No.: 00315 Lin. Date: 04/03/2017 Lin. No.: 00023
Solution File No.: 00322 Soln. Date: 06/18/2017 Soln. No.: 00117
Sequential File No.: 00322 File Date: 06/18/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0096
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 1372

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:11D		
Control 1 EC	0.101%	20:11D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:12D		
Control 2 EC	0.101%	20:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:13D		
Control 3 EC	0.101%	20:14D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 06/18/2017