

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00313	Calib. Date:	04/03/2017
Certification File No.:	00314	Cert. No.:	00024
Linearity File No.:	00315	Cert. Date:	04/03/2017
Solution File No.:	00323	Lin. Date:	04/03/2017
Sequential File No.:	00323	Lin. No.:	00023
		Soln. Date:	07/25/2017
		Soln. No.:	00118
		File Date:	07/25/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZJ-0096
Control Solution %:	0.100%			Expires:	01/30/2019
Solution Control Lot:	17050			Bottle No.:	1363

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:04D		
Control 1 EC	0.097%	22:04D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05D		
Control 2 EC	0.097%	22:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:06D		
Control 3 EC	0.098%	22:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 07/25/2017