

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXD-0004 |
| Location: | SPRING LAKE HEIGHTS PD | |
| Calibration File No.: | 00336 | Calib. Date: 02/26/2018 |
| Certification File No.: | 00337 | Calib. No.: 00026 |
| Linearity File No.: | 00338 | Cert. Date: 02/26/2018 |
| Solution File No.: | 00351 | Cert. No.: 00023 |
| Sequential File No.: | 00351 | Lin. Date: 02/26/2018 |
| | | Lin. No.: 00025 |
| | | Soln. Date: 08/04/2018 |
| | | Soln. No.: 00131 |
| | | File Date: 08/04/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDZJ-0097 |
| Solution Control Lot: | 17420 | Expires: 12/11/2019 |
| | | Bottle No.: 0288 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 17:53D | | |
| Control 1 EC | 0.097% | 17:54D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 17:54D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:55D | | |
| Control 2 EC | 0.096% | 17:55D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 17:55D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:56D | | |
| Control 3 EC | 0.097% | 17:57D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 17:57D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:58D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 08/04/2018