

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00375	Calib. No.: 00029
Certification File No.:	00376	Cert. No.: 00026
Linearity File No.:	00377	Lin. No.: 00028
Solution File No.:	00384	Soln. No.: 00144
Sequential File No.:	00384	
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0345

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:58D		
Control 1 EC	0.101%	11:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 2 EC	0.101%	12:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		
Control 3 EC	0.101%	12:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WILLMS

First Name: CASEY

MI: F

Signature: DSC Casey F Willms #4931

Badge No.: 4931

Date: 08/08/2019