

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00313	Calib. Date: 04/03/2017
Certification File No.:	00314	Calib. No.: 00024
Linearity File No.:	00315	Cert. Date: 04/03/2017
Solution File No.:	00324	Cert. No.: 00021
Sequential File No.:	00324	Lin. Date: 04/03/2017
		Lin. No.: 00023
		Soln. Date: 09/04/2017
		Soln. No.: 00119
		File Date: 09/04/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1385

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:56D		
Control 1 EC	0.100%	21:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58D		
Control 2 EC	0.100%	21:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:59D		
Control 3 EC	0.100%	22:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 09/04/2017