

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00375	Calib. Date: 06/28/2019
Certification File No.:	00376	Calib. No.: 00029
Linearity File No.:	00377	Cert. Date: 06/28/2019
Solution File No.:	00387	Cert. No.: 00026
Sequential File No.:	00387	Lin. Date: 06/28/2019
		Lin. No.: 00028
		Soln. Date: 09/15/2019
		Soln. No.: 00145
		File Date: 09/15/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0171

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:30D		
Control 1 EC	0.100%	20:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31D		
Control 2 EC	0.100%	20:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33D		
Control 3 EC	0.099%	20:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 09/15/2019