

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00352	Calib. Date:	08/23/2018
Certification File No.:	00353	Cert. No.:	00027
Linearity File No.:	00354	Lin. Date:	08/23/2018
Solution File No.:	00356	Cert. No.:	00024
Sequential File No.:	00356	Lin. No.:	00026
		Soln. Date:	09/24/2018
		Soln. No.:	00133
		File Date:	09/24/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0096
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:15D		
Control 1 EC	0.102%	20:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16D		
Control 2 EC	0.102%	20:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	20:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18D		
Control 3 EC	0.102%	20:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	20:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_

Badge No.: 4926

Date: 09/24/2018