

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00326	Calib. Date: 09/18/2017
Certification File No.:	00327	Cert. No.: 00025
Linearity File No.:	00328	Cert. No.: 00022
Solution File No.:	00331	Lin. No.: 00024
Sequential File No.:	00331	Soln. No.: 00121
		File Date: 10/31/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDZJ-0097
Control Solution %:	0.100%		Expires: 01/30/2019
Solution Control Lot:	17050		Bottle No.: 1344

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:25S		
Control 1 EC	0.098%	19:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26S		
Control 2 EC	0.097%	19:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:28S		
Control 3 EC	0.098%	19:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 10/31/2017