

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00326	Calib. Date:	09/18/2017
Certification File No.:	00327	Cert. Date:	09/18/2017
Linearity File No.:	00328	Lin. Date:	09/18/2017
Solution File No.:	00333	Soln. Date:	12/09/2017
Sequential File No.:	00333	File Date:	12/09/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0097
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1258

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:32S		
Control 1 EC	0.099%	22:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33S		
Control 2 EC	0.099%	22:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35S		
Control 3 EC	0.098%	22:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 12/09/2017