

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXD-0004  
Location: SPRING LAKE HEIGHTS PD  
Calibration File No.: 00398 Calib. Date: 04/27/2020 Calib. No.: 00032  
Certification File No.: 00399 Cert. Date: 04/27/2020 Cert. No.: 00028  
Linearity File No.: 00400 Lin. Date: 04/27/2020 Lin. No.: 00030  
Solution File No.: 00403 Soln. Date: 05/10/2020 Soln. No.: 00152  
Sequential File No.: 00403 File Date: 05/10/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0097  
Control Solution %: 0.100% Expires: 10/30/2021  
Solution Control Lot: 19300 Bottle No.: 0788

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:27D		
Control 1 EC	0.097%	22:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28D		
Control 2 EC	0.097%	22:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30D		
Control 3 EC	0.097%	22:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: 

Badge No.: 4926

Date: 05/10/2020