

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00398	Calib. Date:	04/27/2020
Certification File No.:	00399	Cert. Date:	04/27/2020
Linearity File No.:	00400	Lin. Date:	04/27/2020
Solution File No.:	00407	Soln. Date:	06/22/2020
Sequential File No.:	00407	File Date:	06/22/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0097
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	0807

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:51D		
Control 1 EC	0.096%	22:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52D		
Control 2 EC	0.096%	22:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53D		
Control 3 EC	0.096%	22:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 06/22/2020