

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00398 Calib. Date: 04/27/2020 Calib. No.: 00032
Certification File No.: 00399 Cert. Date: 04/27/2020 Cert. No.: 00028
Linearity File No.: 00400 Lin. Date: 04/27/2020 Lin. No.: 00030
Solution File No.: 00410 Soln. Date: 07/15/2020 Soln. No.: 00155
Sequential File No.: 00410 File Date: 07/15/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0097
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0352

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:05D		
Control 1 EC	0.096%	00:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		
Control 2 EC	0.096%	00:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07D		
Control 3 EC	0.097%	00:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 07/15/2020