

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00412 Calib. Date: 07/28/2020 Calib. No.: 00033
Certification File No.: 00413 Cert. Date: 07/28/2020 Cert. No.: 00029
Linearity File No.: 00414 Lin. Date: 07/28/2020 Lin. No.: 00031
Solution File No.: 00430 Soln. Date: 11/08/2020 Soln. No.: 00159
Sequential File No.: 00430 File Date: 11/08/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0097
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0235

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:38S		
Control 1 EC	0.098%	19:38S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:39S		
Control 2 EC	0.098%	19:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:40S		
Control 3 EC	0.099%	19:41S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 11/08/2020