Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD

Calibration File No.:	00412	Calib. Date:	07/28/2020	Calib. No.:	00033
Certification File No.:	00413	Cert. Date:	07/28/2020	Cert. No.:	00029
Linearity File No.:	00414	Lin. Date:	07/28/2020	Lin. No.:	00031
Solution File No.:	00431	Soln. Date:	12/14/2020	Soln. No.:	00160
Sequential File No.:	00431	File Date:	12/14/2020		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZJ-0097
Control Solution %:	0.100%			Expires:	03/30/2022
Solution Control Lot:	20180			Bottle No.:	0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:07S		
Control 1 EC	0.097%	21:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:08S		
Control 2 EC	0.098%	21:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10S		
Control 3 EC	0.097%	21:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: RAMP First Name: ZACHARY MI: M
Signature:
Badge No.: 4926
Date: 12/14/2020