

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00433	Calib. Date: 01/13/2021
Certification File No.:	00434	Calib. No.: 00034
Linearity File No.:	00435	Cert. Date: 01/13/2021
Solution File No.:	00447	Cert. No.: 00030
Sequential File No.:	00447	Lin. Date: 01/13/2021
		Lin. No.: 00032
		Soln. Date: 06/15/2021
		Soln. No.: 00171
		File Date: 06/15/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1256

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:03D		
Control 1 EC	0.098%	18:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		
Control 2 EC	0.097%	18:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		
Control 3 EC	0.097%	18:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_

Badge No.: 4926

Date: 06/15/2021