

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00453	Calib. Date: 07/09/2021
Certification File No.:	00454	Calib. No.: 00035
Linearity File No.:	00455	Cert. Date: 07/09/2021
Solution File No.:	00459	Cert. No.: 00031
Sequential File No.:	00459	Lin. Date: 07/09/2021
		Lin. No.: 00033
		Soln. Date: 08/10/2021
		Soln. No.: 00173
		File Date: 08/10/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0097
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1034

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:14D		
Control 1 EC	0.099%	02:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15D		
Control 2 EC	0.099%	02:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		
Control 3 EC	0.099%	02:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____

Badge No.: 4926

Date: 08/10/2021