

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |            |
|-------------------------|------------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARXD-0004  |
| Location:               | SPRING LAKE HEIGHTS PD |              |            |
| Calibration File No.:   | 00453                  | Calib. Date: | 07/09/2021 |
| Certification File No.: | 00454                  | Cert. No.:   | 00035      |
| Linearity File No.:     | 00455                  | Cert. Date:  | 07/09/2021 |
| Solution File No.:      | 00463                  | Lin. Date:   | 07/09/2021 |
| Sequential File No.:    | 00463                  | Lin. No.:    | 00033      |
|                         |                        | Soln. Date:  | 12/12/2021 |
|                         |                        | Soln. No.:   | 00175      |
|                         |                        | File Date:   | 12/12/2021 |
| <br>                    |                        |              |            |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34      |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDZJ-0097  |
| Solution Control Lot:   | 21090                  | Expires:     | 03/17/2023 |
|                         |                        | Bottle No.:  | 0613       |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 20:01S |                |                     |
| Control 1 EC      | 0.098% | 20:02S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 20:02S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:03S |                |                     |
| Control 2 EC      | 0.098% | 20:04S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 20:04S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:04S |                |                     |
| Control 3 EC      | 0.099% | 20:05S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 20:05S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:06S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: \_\_\_\_\_



Badge No.: 4926

Date: 12/12/2021