

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00464	Calib. Date: 12/28/2021
Certification File No.:	00465	Calib. No.: 00036
Linearity File No.:	00466	Cert. Date: 12/28/2021
Solution File No.:	00468	Cert. No.: 00032
Sequential File No.:	00468	Lin. Date: 12/28/2021
		Lin. No.: 00034
		Soln. Date: 02/06/2022
		Soln. No.: 00177
		File Date: 02/06/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 1047

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:22S		
Control 1 EC	0.096%	20:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	20:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:23S		
Control 2 EC	0.096%	20:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	20:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:25S		
Control 3 EC	0.096%	20:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____

Badge No.: 4926

Date: 02/06/2022