

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0004
Location: SPRING LAKE HEIGHTS PD
Calibration File No.: 00464 Calib. Date: 12/28/2021 Calib. No.: 00036
Certification File No.: 00465 Cert. Date: 12/28/2021 Cert. No.: 00032
Linearity File No.: 00466 Lin. Date: 12/28/2021 Lin. No.: 00034
Solution File No.: 00471 Soln. Date: 03/21/2022 Soln. No.: 00178
Sequential File No.: 00471 File Date: 03/21/2022

Calibrating Unit: P Model No.: CU-34 Serial No.: DDZJ-0096
Control Solution %: 0.100% Expires: 02/17/2024
Solution Control Lot: 22070 Bottle No.: 0424

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:11S		
Control 1 EC	0.099%	21:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:12S		
Control 2 EC	0.098%	21:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:14S		
Control 3 EC	0.098%	21:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WILLMS

First Name: CASEY

MI: F

Signature: DSG Casey Willms #4931

Badge No.: 4931

Date: 03/21/2022