

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0004
Location:	SPRING LAKE HEIGHTS PD		
Calibration File No.:	00464	Calib. Date:	12/28/2021
Certification File No.:	00465	Cert. Date:	12/28/2021
Linearity File No.:	00466	Lin. Date:	12/28/2021
Solution File No.:	00481	Soln. Date:	05/15/2022
Sequential File No.:	00481	File Date:	05/15/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0096
Solution Control Lot:	22070	Expires:	02/17/2024
		Bottle No.:	0306

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:34D		
Control 1 EC	0.098%	20:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:36D		
Control 2 EC	0.098%	20:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:38D		
Control 3 EC	0.099%	20:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DAVIS

First Name: IAN

MI: G

Signature: _____

Badge No.: 4934

Date: 05/15/2022