

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0004
Location:	SPRING LAKE HEIGHTS PD	
Calibration File No.:	00433	Calib. Date: 01/13/2021
Certification File No.:	00434	Calib. No.: 00034
Linearity File No.:	00435	Cert. Date: 01/13/2021
Solution File No.:	00444	Cert. No.: 00030
Sequential File No.:	00444	Lin. Date: 01/13/2021
		Lin. No.: 00032
		Soln. Date: 05/11/2021
		Soln. No.: 00168
		File Date: 05/11/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0096
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 0544

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:10D		
Control 1 EC	0.099%	01:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11D		
Control 2 EC	0.099%	01:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13D		
Control 3 EC	0.098%	01:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RAMP

First Name: ZACHARY

MI: M

Signature: _____



Badge No.: 4926

Date: 05/11/2021